

आई.सी.ए.आर. – केन्द्रीयअन्तर्स्थलीयमात्स्यिकीअनुसंधानसंस्थान बैरकपुर, कोलकाता-700120, पश्चिमबंगाल ICAR -CENTRAL INLAND FISHERIES RESEARCH INSTITUTE Barrackpore, Kolkata -700120, West Bengal

Notice Inviting Quotations

F. No. Capital (EFC Equip. 25-26)-309(1)/2025-Stores

Quotations are invited from the eligible vendors for following item(s):

Items required:

Water Quality Monitoring Sensors

| SI. No. | Sensor Type | Specification | Qty. | Delivery Location |
|------------|--------------------|------------------------------|------|----------------------|
| 1 | Dissolved Oxygen | Type: Optical | 1 | Maithon Dam |
| | Sensor | Range: 0 – 20 mg/l | | |
| | | Mode of output Signal: RS485 | | |
| | | Voltage Range: 6 – 12 Volt | | |
| 2 | Ammoina Sensor | Type: Electrochemical | 1 | |
| | | Range: 0 – 1000 mg/l NH4-N | | |
| | | pH: 0 – 14 pH | | |
| | | Mode of output singal: RS485 | | |
| | | Voltage Range: 6 – 12 Volt | | |
| 3 | Turbidity Sensor | Type: Optical | 1 | |
| | | Range: 0 – 1000 NTU | | |
| | | Mode of Output signal: RS485 | | |
| | | Voltage range: 6 – 12 Volt | | |
| 4 | Chlorophyll Sensor | Type: Optical | 1 | |
| | | Range: 0 – 400 ug/l | | |
| | | Mode of output signal: RS485 | | |
| | | Voltage range: 6 – 12 Volt | | |
| | Warranty: 2 Y | ears | | |

- Payment will be made after delivery. TDS will be deducted as per rule.
- Item should be delivered on or before the stipulated time.
- Quotations should be sent preferably in a sealed envelope within 17.12.2025 to the following address:

The Director,

ICAR – Central Inland Fisheries Research Institute,

Barrackpore, Kolkata - 700120

Email id: storeseccifri@gmail.com

Quotation opening date: 18.12.2025 at 11:00 AM

Date: 11-12-2025